08/06/02 1 of 3

# Summary of SAFARI Gas Instrument Calibrations (UW Flights 1810-1840: 10 Aug-18 Sept 2000)

## TECO 43S Pulsed Flourescence SO<sub>2</sub> analyzer

UW Flights 1810-1814 ( 10 Aug-15 Aug):

Instrument was configured with a 200 ppbv range and 2 minute averaging time.

$$SO_2[ppbv] = 40.5(volts) - 1.6$$
  
standard error  $(\sigma) = 2\%$ 

UW Flights 1815-1840 (17 Aug-18 Sept):

Instrument was configured with a 100 ppbv range and 2 minute averaging time.

$$SO_{2}[ppbv] = 1 + \frac{(180-1)[volts - [3.7 \times 10^{-2} + 7.2 \times 10^{-4}(tstatr) - 4.2 \times 10^{-6}(pstat)]]}{[3.7 + 1.7 \times 10^{-2}(tstatr) + 3.6 \times 10^{-4}(pstat)] - [3.7 \times 10^{-2} + 7.2 \times 10^{-4}(tstatr) - 4.2 \times 10^{-6}(pstat)]}$$

$$standard\ error\ (\sigma) = 7\%$$

where "volts" is the voltage output of the TECO 43S (in volts). where "tstatr" is the static air temperature measured by the reverse flow thermometer (in °C). where "pstat" is the pressure measured by the Rosemount Model 830 BA (in mb).

## TEI Model 49C UV Absorption O<sub>3</sub> analyzer

UW Flights 1810-1840 (10 Aug-18 Sept):

Instrument was configured with a 60 second averaging time, 1000 ppbv range, and automatic pressure and temperature corrections.

$$O_3$$
 [ppbv] = -0.7+94.5(volts)  
standard error ( $\sigma$ ) = 1%

where "volts" is the voltage output of the TEI 49C (in volts).

08/06/02 2 of 3

### LI-COR LI-6262 Infrared Correlation Spectrometer CO<sub>2</sub> analyzer

UW Flights 1810-1840 (10 Aug-18 Sept):

Instrument was configured with 1 second averaging time, a 1 ppmv range, and a metal bellows pump pulling sample air through the instrument.

$$CO_{2}[ppmv] = 1 + \frac{(205-1)[volts - [8.2 \times 10^{-1} + 1.4 \times 10^{-2}(tstatr) - 4.5 \times 10^{-4}(pstat)]]}{[1.4 + 5.6 \times 10^{-3}(tstatr) + 5.9 \times 10^{-7}(pstat)] - [8.2 \times 10^{-1} + 1.4 \times 10^{-2}(tstatr) - 4.5 \times 10^{-4}(pstat)]}$$

$$standard\ error\ (\sigma) = 15\%$$

where "volts" is the voltage output of the LI-COR LI-6262 (in volts). where "tstatr" is the static air temperature measured by the reverse flow thermometer (in °C). where "pstat" is the pressure measured by the Rosemount Model 830 BA (in mb).

#### **TECO Model 48 Infrared Correlation Spectrometer CO analyzer**

UW Flight 1816 (18 Aug)

Instrument range malfunctioned: no measurements

UW Flights 1810-1815 (10 Aug-17 Aug ) and UW Flights 1817-1818 (20 Aug):

Instrument was configured with 2 ppmv range and 10 second averaging time.

$$CO[ppbv] = 362.0(volts)-1.7$$
  
 $standard\ error\ (\sigma) = 3\%$ 

UW Flights 1819-1822 (20 Aug-24 Aug):

Instrument was configured with 2 ppmv range and 10 second averaging time.

$$CO[ppbv] = 5 + [(1424-5)(volts + 1.2 \times 10^{-1})/(5.6 \times 10^{-1} + 1.2 \times 10^{-1})]$$
  
 $standard\ error\ (\sigma) = 7\%$ 

UW Flights 1823-1840 (29 Aug-18 Sept):

Instrument was configured with 2 ppmv range and 10 second averaging time, and the zero was raised to reduce noise at low concentrations.

$$CO[ppbv] = 5 + [1424 - 5)(volt - .4)/(1.3 - .4)]$$
  
 $standard\ error\ (\sigma) = 21\%$ 

where "volts" is the voltage output of the TECO Model 48 (in volts).

08/06/02 3 of 3

## Monitor Labs Model 8840 Chemiluminescence NO/NO<sub>x</sub> analyzer

UW Flights 1810-1821 (10 Aug-23 Aug):

Vacuum pump not operational: no measurements

UW Flights 1822-1840 (24 Aug-18 Sept):

Instrument was configured with 1 ppmv range and 60 second averaging time, and vacuum pump was operational.

$$NO[ppbv] = 1 + \frac{(1004-1)[volts - [8.3+8.3\times10^{-2}(tstatr) - 1.0\times10^{-2}(pstat)]]}{[5.8+1.1\times10^{-2}(tstatr) - 4.1\times^{-3}(pstat)] - [8.3+8.3\times10^{-2}(tstatr) - 1.0\times10^{-2}(pstat)]}$$
 
$$standard\ error\ (\sigma) = 37\%$$

$$NO_{x}[ppbv] = 1 + \frac{(1004-1)[volts - [3.7-2.0x^{-2}(tstatr) - 2.1x10^{-3}(pstat)]]}{[8.2+2.5x10^{-2}(tstatr) + 5.6x10^{-3}(pstat)] - [3.7+2.0x10^{-2}(tstatr) - 2.1x10^{-3}(pstat)]}$$
 
$$standard\ error\ (\sigma) = 30\%$$

where "volts" is the voltage output of the Monitor Labs Model 8840 (in volts). where "tstatr" is the static air temperature measured by the reverse flow thermometer (in °C). where "pstat" is the pressure measured by the Rosemount Model 830 BA (in mb).

\_\_\_\_\_